

Servo controlled Infant Care Incubator Mounted on Trolley Model: Neo 1001

General Specifications:

- Clear Acrylic Wall Canopy with four port holes
 Front Loading
- Four Inlets for administration of O2, IV lines or Ventilator tubing.
- Sliding Baby Tray with External jerk free gear based Head Up / Down facility
- Stainless Steel rust free Heater, blower sub assembly with changeable micro Air filters.
- Humidity heater and Hygrometer.
- Pole for Syringe / Infusion Pump & S.S. IV Stand
- · Lower two trays for utility
- Trolley mounted on four anti-skid castors with locking facility

Technical Specifications:

- Modes SKIN, AIR and Manual Control
- Temperature setting from 19°C to 39°C with heater power bar graph indication
- Display two bright LED digital displays combination of Red and Green to continues monitor of baby temp and set temp, with easily switchable to °C to °F
- Safety independent Hi temp cut off with Hi temp sensor at 39°C
- Power AC 240V, 50 Hz, 1000W Maximum, Fuse rating 6 Amps
- Alarms both audio and Visual indications for Baby Probe Failure,
 Air Probe Failure,
 Heater Failure,
 1°C high,1°C low,
 Hi temp cut off,
 System failure,
 Power fail indication.

Model: Neo 1001

Note: Technical specifications may be altered by NEOKRAFT without prior notice, Accessories shown in the catalogue are not part of the standard equipment



Neokraft Medical Pvt Ltd

#. 18, 1st Cross, 11th Main, Tata Nagar, Sahakara Nagar Post,

BANGALORE-560 092, INDIA.

Phone: 080 2341 5855 / 080 6500 0249

Cell: +91 9686 78 5555

e mail: info@neokraftmedical.com web: neokraftmedical.com